



## Addendum for FOX MOBILE Inhalation System IFU DAW-00045 (rev 6)

### 9.2 Aerosol characteristics

The Mass Median Aerosol Diameter (MMAD) values presented in section 9.2 (Page 36) are incorrect.

The correct MMAD values are shown below:

PARTICLE SPECIFICATION	SULTANOL	CROMOHEXAL	ATROVENT
Mass Median Aerosol Diameter (MMAD)	4.0 $\mu\text{m} \pm 0.1$	4.0 $\mu\text{m} \pm 0.1$	4.0 $\mu\text{m} \pm 0.1$